

Glow Wire Tester Model 4180

Friborg Test Technology AB

Mail address

P.O. Box 63
SE-196 22 Kungsängen
Sweden

Visiting address

Kraftvägen 32
SE-196 37 Kungsängen
Sweden

T: +46 (0) 8 581 610 10
F: +46 (0) 8 505 961 50

info@friborg.se
www.friborg.se

Technical specifications

- The equipment is fully automatic
- PLC is controlling moving of carriage, time Intervals and burning time for sample
- Flame sensor indicates start/stop burning
- Digital display presets testing alternatives
- 16 test results are memorized in the PLC
- Optional for printer output or memorized test results to PC
- Remote control outside the cabinet operates the equipment when doors are closed
- The model 4180 means effective testing and documentation of Glow Wire Test results

Optional – Cabinet for Glow Wire

- Cabinet 0,5 m³ according to requirement of standard
- 1 pcs transparent door
- Black painted inside of cabinet for visual inspection of flames
- Weight 60 kg
- LxWxH = 1130x780x1000 mm
- Fan blow out gases automatic (if required) from the cabinet after testing

Facts

IEC:

60695-2-10, 60695-2-11,
60695-2-12 and 60695-2-13

Components:

-

Weight:

App 40 kg

Dimensions:

App 940x450x400 mm
(LxWxH) without cabinet

Supply

230/115 VAC, 50/60 Hz,
3.15 A / 6.3 A

